
Prostaglandin E2 CLIA Kit

Catalog # KA0324 Size 1 Kit

Applications



The standard curve is for the purpose of illustration only and should not be used to calculate unknowns. A standard curve should be generated each time the assay is performed.

Specification

Product Description	Prostaglandin E2 CLIA Kit is a competitive immunoassay for the quantitative sensitive determination of Prostaglandin E2 in biological fluids.
Suitable Sample	Saliva, Serum, Tissue Culture Supernatant, Urine, Whole Blood
Sample Volume	100 uL
Label	AP-conjugate
Detection Method	Colorimetric
Assay Type	Quantitative
Spiking Recovery	Cell Culture Supernatant: 102.40%, Male Serum: 126.10%, Female Serum: 113.70%, Saliva: 123.30 %, Urine: 108.90%, Whole Blood: 113.60%
Calibration Range	7.81 to 1000 pg/mL
Limit of Detection	6.03 pg/mL
Regulation Status	For research use only (RUO)
Quality Control Testing	Standard curve The standard curve is for the purpose of illustration only and should not be used to calculate unknown s. A standard curve should be generated each time the assay is performed.

Copyright © 2023 Abnova Corporation. All Rights Reserved.

Storage Instruction	The standard and conjugate must be stored at -20°C; others store at 4°C.

Applications

- Quantification

Copyright © 2023 Abnova Corporation. All Rights Reserved.