
CytoQuest[™] Colorectal Cancer EpCAM PanCK CD45 CDX2 Antibody Kit

Catalog # KA4785 Size 1 Kit

Applications



Immunofluorescence (Circulating Colorectal Cancer Cell)

Representative images of CTC (white arrow) and WBC (yellow arrow) from colorectal cancer patient. CTC was detected by using immunofluorescence staining for PanCK (FITC, green), CDX2 (Alexa 647, red), CD45 (PE, orange) and Nucleus (Hoechst, blue).

| Specification | |
|------------------------|--|
| Product Description | CytoQuest™ Colorectal Cancer EpCAM PanCK CD45 CDX2 Antibody Kit contains antibodies for i mmobilization and immunostaining of circulating colorectal cancer cells. |
| Instrument Requirement | <u>CytoQuest™ CR</u> |
| Chip Requirement | <u>CytoChipNano</u> |



Product Information

| Supplied Product | Kit content: 1. Anti-EpCAM capturing antibody (Biotin): Biotin conjugated Anti-EpCAM antibody for circulating colorectal cancer cell capturing. |
|---------------------|---|
| | 2. Anti-PanCK detecting antibody (FITC): FITC conjugated Anti-PanCK antibody for circulating colorectal cancer cell detection. |
| | 3. Anti-CDX2 detecting antibody : Anti-CDX2 antibody for circulating colorectal cancer cell detection. |
| | 4. Anti-CD45 detecting antibody (PE): PE conjugated Anti-CD45 antibody for circulating colorectal cancer cell detection. |
| | 5. Secondary antibody (Alexa 647) |
| | 6. 50X Antibody Dilution Buffer (50X ADB). |
| | *Reagents are sufficient for 20 assays using recommended protocol. |
| Regulatory Status | For research use only (RUO) |
| Storage Instruction | Store Anti-EpCAM capturing antibody (Biotin), Anti-PanCK detecting antibody (FITC), Anti-CDX2 de tecting antibody (Alexa 647), Anti-CD45 detecting antibody (PE) and Secondary antibody (Alexa 647) at 4°C. |
| | Store Anti-CDX2 detecting antibody and 50X Antibody Dilution Buffer (50X ADB) at -20°C. |
| | Aliquot to avoid repeated freezing and thawing. |

Applications

Immunofluorescence (Circulating Colorectal Cancer Cell)

Representative images of CTC (white arrow) and WBC (yellow arrow) from colorectal cancer patient. CTC was detected by using immunofluorescence staining for PanCK (FITC, green), CDX2 (Alexa 647, red), CD45 (PE, orange) and Nucleus (Hoechst, blue).